

Use, Disclaimer, and Warnings

The P4L HomeCLINIC creates a powerful electrical magnetic field that extends approximately 3 feet in circumference from the device applicators. These powerful therapeutic magnetic pulses may erase some magnetic strips on credit cards and identifications cards, as well as damage or disrupt electronics. It is important to keep the following items at least 2 feet from the device applicators when in use:

- Credit cards
- Cell phones
- Identification cards or tags with magnetic strips
- Car keys with electronic key fobs
- Computers & other electrical devices

NEVER USE THE P4L HOMECLINIC DEVICE IN, NEAR, OR AROUND WATER

The P4L HomeClinic powerful magnetic pulses **prohibit** use under the following circumstances:

- During pregnancy
- Those with non-MRI safe IUD
- People or animals with organ transplants
- People or animals with pacemakers, hearing aids, medical pumps, or other electrical implants/devices

NOTE: Non-magnetic metal rods, screws, and plates are safe for PEMF use. If an implant is safe for an MRI, it is safe to use the A2000.

If you have any concerns regarding your pet's microchip, two of the largest animal ID microchip manufacturers in the U.S. state that their product has been rated for electromagnetic exposure up to 7 Tesla. This is well beyond the magnetic strength of the P4L HomeCLINIC. We highly encourage pet owners to contact your pet's ID microchip supplier, if you have additional concerns.

USE: Remove metal chains, collars, and items that may contain magnetic strips, or that may be electronically-operated.

If the user is in frail condition, limit early exposure to 15 minutes per treatment for the first few days. Therapy session times may be increased, according to user's preference.

Pulsed electromagnetic fields (PEMF) supports the natural elimination of toxins. While this elimination is beneficial, it is very important that humans and/or animals being treated, drink a sufficient amount of water to assist this process.